



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NCSL Z540-1-1994

HOFFER CALIBRATION SERVICES, LLC
 107 Kitty Hawk Lane
 Elizabeth City, NC 27909
 Wendy Brabble Phone: 252 331 1997

CALIBRATION

Valid to: April 30, 2020

Certificate Number: 4916.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Fluid Quantities

Parameter/Equipment	Range	CMC ² (±)	Comments
Volumetric Flow Rate – Light Hydrocarbons	(0.2 to 60) gpm	0.05 %	FDI-PDCL-60
	(0.757 to 227.124) lpm		
	(0.2 to 400) gpm	0.05 %	Omnitrak OT-400
	(0.757 to 1514.16) lpm		
	(0.2 to 60) gpm	0.05 %	FDI-PDCL-60
	(0.757 to 227.124) lpm		
Water	(0.01 to 10) gpm	0.05 %	MT-10
	(0.0378 to 37.854) lpm		
	(1.0 to 260) gpm	0.20 %	Transfer standard
	(3.785 to 984.207) lpm		
	(0.75 to 243) gpm	0.20 %	
(2.839 to 919.855) lpm			
Air	(5.0 to 1525) gpm	0.20 %	
	(18.927 to 5772.753) lpm		
	(0.03 to 200) acfm	0.30 %	CFSNC-250

Parameter/Equipment	Range	CMC ² (±)	Comments
Volumetric Flow Rate ³ – Cryogenic–Liquid Nitrogen	(20 to 60) gpm	0.60 %	Prover 1"
	(75.708 to 227.124) lpm		
	(20 to 130) gpm	0.60 %	Prover 1 ^{1/2} "
	(75.708 to 492.102) lpm		
	(20 to 200) gpm	0.60 %	Prover 2"
	(75.708 to 757.082) lpm		
Cryogenic-Liquid Carbon Dioxide (CO ₂)	(20 to 60) gpm	0.40 %	Prover 1"
	(75.708 to 227.124) lpm		
	(20 to 130) gpm	0.40 %	Prover 1 ^{1/2} "
	(75.708 to 492.102) lpm		
	(20 to 200) gpm	0.40 %	Prover 2"
	(75.708 to 757.082) lpm		

¹ This laboratory offers commercial calibration service and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.